



# CERTIFICATE OF CALIBRATION

Dwyer Instruments, 102 Highway 212, Michigan City, IN 46360 USA  
 T: +1 800.872.9141 +1 219.879.8000 F: +1 219.872.9057

ID/Serial#: 00VVD7		Date: 2/24/2017	Date Due: N/A	Bv:
<b>Customer Information</b>		<b>Device Under Test (DUT)</b>		
		ID/Serial#:	00VVD7	
		Type:	Series 2000 Magnehelic	
		Description:	Pressure Measurement Device	
		MFR:	Dwyer Instruments	
		Model:	2000-60PA - HA	
		Accuracy:	+/-2%FS	
		Variable:	Pressure	Units: Pa
<b>Address Where Calibration Was Performed</b>		<b>Calibration Procedure</b>		
Dwyer Instruments 55 Ward St Wakarusa, IN 46573		Procedure:		
		Reference:		
		Status:	As Found/As Left	
		Test Range:	0	To 60 Pa
		Notes:		
<b>Reference Standards Used</b>				
Module	ID#	Last Cal	Due Date	
GQS6180		3-1-16	3-31-17	
	MFC18	10-14-16	10-14-17	

As Found/As Left Calibration Data, Pressure (absolute) (Pa)				
RESULT	DUT Target	Test Point	Reference Standard	% Error Full Scale
PASSED	0		0	0
PASSED	12		12	0
PASSED	24		24.05	-0.08
PASSED	36		36.27	-0.45
PASSED	48		48.24	-0.4
PASSED	60		59.96	0.07

Signature: Lynn Deckmann Job Function: Line Leader